

ANNEX A – TECHNICAL SPECIFICATION INVITATION TO BID: No. ITB/HCR/KSO/2022/001

PROVISION FOR SUPPLY AND DELIVERY OF MEDICAL EQUIPMENT TO UNHCR OFFICE KASSALA.

CLOSING DATE AND TIME: 22 NOVEMBER 2022 - 23:59 HRS.

Supply and delivery of (MEDICAL EQUIPMENT) to UNHCR in Kassala Office – Sudan

S. No.		Item descriptions	
1	Blood Hematology analyzer		
	Specifications	CBC+3-DIFF, 19parameters+3histograms Throughput:40 samples per hour Intuitive operation system with 8.4 inch TFT touchscreen Open vial sampling 200,000 results storage with histograms Original QC, calibrators, and reagents	
2	Diagnostic Imaging Machine		
	Specifications	Frequency: 60 KHZ Phase: Single Phase Peak Kilo Voltage: 125 kvP Technology Type: Digital Machine Frequency Type: High Frequency Application: Hospital Output: 500 mA Voltage: 230 V Features:	
3	•Upper and lower extremities Sphygmomanometer mercurial model		
	Specifications	Usage/Application: Hospital Cuff Type: Rubber Scale: 0 To 300 Mm Of Hg Purity of Mercury: 99.99% Display: Mercurial Pack Type: Box	
4	<u>Stethoscope</u>		
	Specifications	Applications Cardiology/High Performance Chestpiece Finish Polished Stainless Steel Chestpiece Technology Single sided Diaphragm Diameter 1.75 Inch Diaphragm Diameter (metric) 4.4 cm	



		Diaphragm Material Epoxy/Fiberglass		
		Diaphragm Type Tunable Diaphragm		
		Eartip Type Soft Sealing		
		Extra Eartips Yes		
		Headset Material Wide-diameter aerospace alloy / Anodized		
		aluminum		
		Length 27 Inch		
		Length (Metric) 69 cm		
		Net Weight 185 g		
		Net Weight (Metric) 185 g		
		Occupation Anesthesiologist, Cardiologist, Emergency Physician,		
		EMT/EMS, Family Practitioner, Internist, Medical Student, Nurse,		
		Pediatrician, Physician, Respiratory Specialist		
		Patient Adult, Infant, Pediatric		
		Performance 10		
		Special Adaptors Yes		
5	Oxygen regulator			
	Specifications	Pressure : 50 psi		
		Gauge: 0 - 4000 psi		
		Connector : CGA 540		
		Flowmeter: 0 - 3 L/min (60 L/min Max) or		
		0 - 5 L/min (60 L/min Max) or		
		0 - 15 L/min (60 L/min Max)		
		with D.I.S.S. Connector		
Note: Please provide the details about the brand, country of origin, warranty period and model				

Note: Please provide the details about the brand, country of origin, warranty period and model of each medical equipment and should be part of the technical offer.

Company Name	
Company Stamp	
Date:	