

#### ANNEX B – FINANCIAL OFFER FORM

#### FOR THE PROVISION OF MEDICAL EQUIPMENT TO UNHCR EAST DARFUR, EL DAEIN HOSPITAL

No	Description	Technical Specifications & FunctionalRequirements	Sample Photo	UoM	QTY	UNIT PRICE DAP UNHCR El Daein	TOTAL AMT DAP UNHCR El Daein
1	A Complete Blood Count (CBC): Sysmex XW-100.	<ul> <li>Compact machine</li> <li>A height of 13.8 inches.</li> <li>A width of 7.3 inches. can fit on a countertop.</li> <li>The Sysmex XW-100 and its reagents</li> </ul>		1			
2	Electrocardiograph (ECG) Zou Care Machine	<ol> <li>Size &amp; Weight:         <ul> <li>Sturdy &amp; light weight machine &lt;5kg,</li> <li>Should be compact</li> <li>Should have carry handle for portability</li> </ul> </li> <li>Power Supply:         <ul> <li>Compatibility with mains 220-240 V (normal), 50/60 Hzpower supply</li> <li>High performance Li-ion rechargeable battery with built-in chargerEquipment should have sufficient battery backup for taking minimum 100 ECGs without AC power.</li> <li>Digital filtering to remove interference from power line,muscle tremor etc.</li> </ul> </li> <li>ECG recording:         <ul> <li>ECG recording with 12 leads</li> <li>ECG recording with 12 leads</li> <li>Standard Leads (the limb leads or bipolar Limb leads: I, II &amp; III)</li> <li>Augmented Limb Leads- (aVL, aVR and aVF)</li> <li>Chest Leads (the unipolar or V leads)- from V<sub>1</sub> to V<sub>6</sub>Simultaneous acquisition from</li> </ul> </li> </ol>		1			



12 leads.
<ul> <li>Recording speed selection of 25 mm/ sec and 50 mm/ secwith facility for speed</li> </ul>
selection
Automatic adjustment of baseline for optimal recording Should have different filters like     Baseline Filter, EMG Filter &AC Filter
Multiple operating modes -automatic, manual and rhythm
Common Mode Rejection Ratio >90dB.
4- Built-in ECG Parameters measurement and interpretation:
Built-in ECG auto-measurement including HR, P-R interval, P-Duration, QRS duration, Q-T
interval, Q-TcF (Fridericia), PAxis, QRS Axis, T Axis, R(V5), S(V1), R(V5) +S(V1)
QTcF interval reading/ measurement should also beavailable with Limb leads alone.
5- Printing and Communication:
High-resolution thermal printing array system
Built in printer should work with standard universal thermalprinter paper.
6- Standard Accessories:
The machine should be supplied with
Power cord, patient cable, user manual and warranty card,
Operation Manual with user demonstration video CD, interpretation manual &10
thermal recording paper rolls, 5 bottles of jelly.
Two sets each of:
Patient cable
Chest electrodes – Both adult and pediatric (2 setseach)
• Limb electrodes – Both adult and pediatric (2 setseach)
7- Safety Profile
Should be provided with terminal for good earth connection preclude electric
disturbances while recording-
Must have a safety certificate or valid detailed electrical andfunctional safety test report
from a recognized competent authority
Copy of the certificate / test report shall be produced alongwith the technical bid.



		<ul> <li>8- Installation &amp;Training:         <ul> <li>The firm should install the instrument at the designated location and provide one-day training/ demonstration of operation of ECG machine</li> <li>9- Warranty:                 <ul> <li>Performance warranty of one year from date of installation (comprehensive warranty)</li> <li>In case of breakdown of the machine, the supplier shall make the machine functional by repair (including replacement of parts/replace the machine if necessary) free of cost at the user site, within 7 §even) days of the receipt of complaint/report.</li></ul></li></ul></li></ul>			
3	Blood Bank Refrigerator –	BRAND NAME: HAIER Inner Temperature: 2-6 C Effective volume: 358 L Rated voltage: 220-244 V Rated Frequency: 50Hz/60Hz Power input: 460 W Rated current: 3 A Power of the lamp: 3 W Climate type: ST CO2, Equivalent (Tones): 0.601 Net weight: 165 kg Refrigerant R134a(GQP:1430): 420 kg Anti-shock safety classification: 1 Power connection type: Y	1		



#### OTHER INFORMATION PARTERNING TO YOUR QUOTATION

Other Information pertaining to our RFQ are as follows:	Yes, we will Comply State, as required	No, we cannot comply (State as required)
Confirmation that your goods meet UNHCR specifications: UNHCR shall not issue a partial payment or cancel part of the order because the contractor is not able deliver products as per specifications. By submitting an offer, you accept that UNHCR should withhold 100% of payment until when you would have delivered the right products in full.		
Offer Validity: (minimum 30 Calendar Days minimum).		
Confirmation that your financial offer has no cancelations or corrective fluid. (Any corrective fluid on your offer will lead to disqualification)		
Acceptance of UNHCR General Terms and conditions of Contract for Provision of Goods		
What is your delivery lead time in days? (Maximum 21 days)		
What is the currency of your offer?		
Name of your company		
Company email		
Company telephone number (two numbers)		
Name of authorized person		
Signature, stamp and date		