# BILL OF QUANTITIES - BOQ 1.4

## Supply, Installation, Preparation, and Training Use for the OBS/GYN Outpatient Clinic - Al Ban Jadeed Teaching Hospital

| Item No. | Item | Description / technical specification | Country of Origin | Photos | Warranty | Installation (YES /NO ) | Training (YES /NO ) | Lead Time |
|---|---|---|---|---|---|---|---|---|
| 1.4.1 | **Portable ultrasound (obstetric and transvaginal probes)** | **Supply and installation of Portable ultrasound, B/M modes (colour & PW Doppler preferred/optional), convex 2-5 MHz + transvaginal 4-9 MHz probes, >= 2 ports, full OB/GYN calcs & reports, >= 15" display, HDD/SSD, USB/DICOM, battery >= 60 min; trolley, B/W printer, gel, TV covers.** | | | | | | |
| 1.4.2 | **Gynaecological examination couch** | **Supply and installation of a Gynae examination couch, steel frame, washable vinyl upholstery, adjustable backrest 0-70 deg, leg supports/stirrups, SS drainage tray, paper roll holder, SWL >= 150 kg, with a 2-step stool** | | | | | | |
| 1.4.3 | **Speculum and gynae instrument sets** | **Supply of Gynae instrument lot : Cusco specula S x4/M x6/L x4, Sims x2, sponge forceps x4, tenaculum x2, uterine sound x2, IUD forceps x2, long scissors x2, kidney dishes x4, gallipots x4, trays x2; I** | | | | | | |
| 1.4.4 | **Examination lamp** | **Supply and installation of a mobile LED examination light, >= 30,000 lux at 0.5 m, 4000-5000 K, CRI >= 90, dimmable, gooseneck/articulated arm, 5-castor base, battery option preferred;** | | | | | | |
| 1.4.5 | **BP apparatus and adult scale** | **Supply of 1 aneroid BP set 0-300 mmHg (+/-3), adult & large cuffs, dual-head stethoscope + 1 adult floor scale >= 150 kg, class III; CE;** | | | | | | |

Company Name:

Company Stamp:

Date: